# Bard® Intermittent Catheters (Latex)

#### Instructions For Use

Indications For Use: For urological use only.



Single-use. Single patient use, do not reuse.



) Do not use if package is opened or damaged.

Storage: Store catheters at room temperature away from direct exposure to light, preferably in the original box.



Caution: This product contains natural rubber latex which may cause allergic reactions.

## Warnings:

- On latex and red rubber catheters, do not use ointments or lubricants having a petrolatum base. They will damage latex.
- · Visually inspect the product for any imperfections or surface deterioration prior to use.
- Reuse and/or repackaging may create a risk of patient or user infection, compromise the structural integrity and/or
  essential material and design characteristics of the device, which may lead to failure, and/or to injury, illness or
  death of the patient.



**Note:** After use, this product may be a potential biohazard. Handle and dispose of in accordance with accepted medical practice and applicable local, state and federal laws and regulations.

## **Adverse Reactions**

Urinary Tract Infection Bleeding from the urethra Irritation of the urethra

### Instructions for Intermittent Catheters:

- Wash your hands thoroughly with soap and water.
- 2. Remove catheter from the pack.
- 3. Position yourself comfortably, cleaning the opening of the urethra and surrounding area.
- 4. Gently insert rounded end of catheter into urethra.
- 5. When urine stops flowing, remove catheter from urethra.
- 6. Dispose of catheter in accordance with local rules and regulations.
- Wash your hands.

R<sub>Only</sub> Caution: Federal (U.S.A.) laws restricts this device to sale by or on the order of a physician.





Manufacturer:

C. R. Bard, Inc.

Covington, GA 30014

1-800-526-4455

www.bardmedical.com

Assembled in Mexico